

Quantum Imaging and Therapeutic Associates

Elbow (Biceps) – 1.5 Tesla

Indications	Biceps tendon tear
Coil	Knee, flex M
Position	Prone, arm above head, thumb up; or Supine, arm at side, palm up; or May flex elbow 90 degrees if patient cannot tolerate
FOV	16 cm
Comments / Special	Must include radial tuberosity distally on Ax Must include biceps tendon anteriorly on Cor/Sag

<u>Sequence</u>	<u>TR</u>	<u>TE</u>	<u>TI</u>	<u>Flip</u>	<u>ST/sp</u>	<u>Matrix</u>
Ax T1	400-800	Min.			3/1	256 x 256
Ax PD FS	3000	30-40			3/1	256 x 256
Cor PD FS	3000	30-40			3/1	256 x 256
Sag T1	400-800	Min.			3/1	256 x 256
Sag T2 FS	>2000	50-60			3/1	256 x 256