

Quantum Imaging and Therapeutic Associates

Finger – 1.5 Tesla

Indications	Fracture, tendon pathology, ligament/pulley tear
Coil	Flex S
Position	Palm down, fingers extended
FOV	10 cm
Comments / Special	N/A

<u>Sequence</u>	<u>TR</u>	<u>TE</u>	<u>TI</u>	<u>Flip</u>	<u>ST/sp</u>	<u>Matrix</u>
Ax T1	400-800	Min.			3/0.3	256 x 192
Ax T2 FS	>2000	50-60			3/0.3	256 x 256
Cor T1	400-800	Min.			3/0.3	256 x 192
Cor PD FS	>3000	30-40			3/0.3	256 x 256
Cor STIR	>2000	20-40	150		3/0.3	256 x 192
Sag PD FS	>3000	30-40			3/0.3	256 x 256